

MDR Database - MDRQ #1706

MDRQ # 1690 (Released): ANNEX I GENERAL SAFETY AND PERFORMANCE REQUIREMENTS

MDRQ # 1691 (Released): CHAPTER I GENERAL REQUIREMENTS

MDRQ # 1704 (Released): 4. Risk control measures adopted by manufacturers for the design and manufacture of the devices shall conform to safety principles, taking account of the generally acknowledged state of the art.

In selecting the most appropriate solutions, manufacturers shall, in the following order of priority:

11/04/2017 02:17 PM - Anonymous

Status:	Released	
Priority:	Low	
Description		
In selecting the most appropriate solutions, manufacturers shall, in the following order of priority:		
Subissues:		
MDRQ # 1707: (a) eliminate or reduce risks as far as possible through safe design and m...		Released
MDRQ # 1708: (b) where appropriate, take adequate protection measures, including alarms...		Released
MDRQ # 1709: (c) provide information for safety (warnings/precautions/contra-indication...		Released